

M-40-1 流量计测试报告

M-40-1 Inspection Report of Flowmeter

产品编号 Number: 22011013

日期 Date: 2022-02-16

流速 Flowrate (L/min)	预定值 Scheduled Volume(L)	游标值 Vernier Show(mm)	真实值 Actual Volume(L)	误差 Error (Ev%)	重复性 Repetitive Error(L%)	备注 Remark
250	501.38	179.00	501.30	+0.016	0.004	
	501.50	182.00	501.40	+0.02		
100	500.85	140.00	500.00	+0.17	0.01	
	500.80	140.00	500.00	+0.16		
25	100.00	129.00	100.00	0	0	
	100.00	129.00	100.00	0		
				1.7%		

装配 Assy: 郭凤英

测试员 Test: 钟发军



质检员 Check: 王卫彬